

Serial #11425

	Nominal	Measured	Adjusted	Error
Date		Sep 22, 2013		
Temperature		21.0 °C		
Humidity		65%		
Units	0.00	0.024	0.012	0.012
	1.00	1.025	1.001	0.10%
	2.00	2.027	2.003	0.15%
	3.00	3.026	3.002	0.07%
	4.00	4.026	4.002	0.05%
	5.00	5.026	5.002	0.04%
	6.00	6.028	6.004	0.07%
	7.00	7.027	7.003	0.04%
	8.00	8.026	8.002	0.03%
	9.00	9.028	9.004	0.04%
	10.00	10.029	10.005	0.05%
Tens	10.00	10.026	10.002	0.02%
	20.00	20.028	20.004	0.02%
	30.00	30.030	30.006	0.02%
	40.00	40.035	40.011	0.03%
	50.00	50.039	50.015	0.03%
	60.00	60.062	60.038	0.06%
	70.00	70.063	70.039	0.06%
	80.00	80.065	80.041	0.05%
	90.00	90.072	90.048	0.05%
	100.00	100.072	100.048	0.05%
Hundreds	100.00	100.000	99.976	-0.02%
	200.00	199.955	199.931	-0.03%
	300.00	299.908	299.884	-0.04%
	400.00	399.890	399.866	-0.03%
	500.00	499.900	499.876	-0.02%
	600.00	599.920	599.896	-0.02%
	700.00	699.950	699.926	-0.01%
	800.00	799.910	799.886	-0.01%
	900.00	899.890	899.866	-0.01%
	1000.00	999.920	999.896	-0.01%
Thousand	1000.00	999.810	999.786	-0.02%
	2000.00	1999.390	1999.366	-0.03%
	3000.00	2998.93	2998.906	-0.04%
	4000.00	3998.60	3998.576	-0.04%
	5000.00	4998.20	4998.176	-0.04%
	6000.00	5997.80	5997.776	-0.04%
	7000.00	6997.30	6997.276	-0.04%
	8000.00	7997.00	7996.976	-0.04%
	9000.00	8996.90	8996.876	-0.03%
	10000.00	9996.20	9996.176	-0.04%